



AssayMax Human Apolipoprotein C-I ELISA Kit

Catalog No. EA8011-1

Introduction

Apolipoprotein C-I (apoC-I) is a 6.6 kDa apolipoprotein that is expressed primarily in the liver and activated when monocytes differentiate into macrophages. After being synthesized as a precursor with a length of 83 amino acids, apoC-I is processed to a single chain mature protein of 57 amino acids (1). It circulates in plasma and is a component of VLDL, IDL, and HDL (2 - 3). ApoC-I plays important modulatory roles in lipoprotein metabolism. It is an inhibitor of lipoprotein binding to the LDL receptor, LDL receptor-related protein, and VLDL receptor (4 - 5). It is the major plasma inhibitor of cholesteryl ester transfer protein and appears to interfere directly with fatty acid uptake (6 - 7). ApoC-I causes hypertriglyceridemia by inhibition of the lipoprotein lipase-dependent triglyceride-hydrolysis pathway (8). On the other hand, apoC-I is an activator of lecithin cholesterol acyl transferase that esterifies cholesterol and produces the formation of the mature HDL (9 - 10). It is also a physiological protector against infection by enhancing the early inflammatory response to lipopolysaccharide (11).

Principal of the Assay

The AssayMax Human Apo C-I ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of human Apo C-I in plasma, serum and cell culture samples. This assay employs a quantitative sandwich enzyme immunoassay technique that measures human Apo C-I in less than 5 hours. A polyclonal antibody specific for human Apo C-I has been pre-coated onto a 96-well microplate with removable strips. Apo C-I in standards and samples is sandwiched by the immobilized antibody and biotinylated polyclonal antibody specific for Apo C-I, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution

Reagents

- **Human Apo C-I Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against human Apo C-I.

- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Human Apo C-I Standard:** Human Apo C-I in a buffered protein base (2 µg, lyophilized, Store at -20°C)
- **Biotinylated Apo C-I Antibody (25x):** A 25-fold concentrated biotinylated polyclonal antibody against Apo C-I (300 µl).
- **EIA Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).
- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (80 µl).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

Storage Condition

- Store kit at 2-8°C or -20°C upon arrival up to the expiration date.
- Opened EIA Diluent may be stored for up to 1 month at 2-8°C. Store reconstituted reagents at -20°C or below.
- Opened unused strip wells may return to the foil pouch with the desiccant pack, reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes (1-20 µl, 20-200 µl, 200-1000µl and multiple channel)
- Deionized or distilled reagent grade water

Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2000 x g for 10 minutes and assay. Dilute samples 1:400 into EIA Diluent as follows: add 10 µl of sample to 390 µl of EIA Diluent (1:40) to make Solution A; then add 80 µl of Solution A to 720 µl of EIA Diluent (1:10) to make a final working solution (1:400). Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles. (EDTA or Heparin can also be used as anticoagulant.)
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2000 x g for 10 minutes. Dilute samples 1:400 into EIA Diluent as follows: add 10 µl of sample to 390 µl of EIA Diluent (1:40) to make Solution A; then add 80 µl of Solution A to 720 µl of EIA Diluent (1:10) to make a final working solution (1:400). Store samples at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Cell Culture Supernatants:** Centrifuge cell culture media at 2000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20°C or below. Avoid repeated freeze-thaw cycles.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use. If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved.
- **EIA Diluent Concentrate (10x):** Dilute the EIA Diluent 1:10 with reagent grade water. Store for up to 1 month at 2-8^oC.
- **Standard Curve:** Reconstitute the 2 µg of Apo C-I Standard with 1 ml of EIA Diluent to generate a solution of 2 µg/ml. Allow the standard to sit on ice for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the standard solution (2 µg /ml) 1:2 with EIA Diluent to produce 1, 0.5, 0.25, 0.125 and 0.063 µg/ml solutions. EIA Diluent serves as the zero standard (0 µg/ml). Any remaining solution should be frozen at -20^oC.

Standard Point	Dilution	[Apo C-I] (µg/ml)
P1	Standard (2 µg/ml)	2.000
P2	1 part P1 + 1 part EIA Diluent	1.000
P3	1 part P2 + 1 part EIA Diluent	0.500
P4	1 part P3 + 1 part EIA Diluent	0.250
P5	1 part P4 + 1 part EIA Diluent	0.125
P6	1 part P5 + 1 part EIA Diluent	0.063
P7	EIA Diluent	0.000

- **Biotin Apo C-I Antibody (25x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:25 with EIA Diluent. Any remaining solution should be frozen at -20^oC.
- **Wash Buffer Concentrate (20x):** Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with EIA Diluent. Any remaining solution should be frozen at -20^oC.

Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30^oC).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Apo C-I standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.
- Wash five times with 200 µl of Wash Buffer manually. Invert the plate each time and decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid. If using a machine wash six times with 300 µl of Wash Buffer and then invert the plate, decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid.
- Add 50 µl of Biotinylated Apo C-I Antibody to each well and incubate for one hour.
- Wash a microplate as described above.
- Add 50 µl of Streptavidin-Peroxidase Conjugate to each well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.


- Wash a microplate as described above.
- Add 50 μ l of Chromogen Substrate per well and incubate for about 30 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 μ l of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a Standard Curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



Performance Characteristics

- The minimum detectable dose of Apo C-I is typically 60 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.3% and 7.4% respectively.
- No cross-reactivity was observed with human Apo C-II at 10 μ g/ml.
- No cross-reactivity was observed with human Apo C-III at 2 μ g/ml.
- No significant cross-reactivity was observed with human Apo A-I, Apo A-II, Apo B and Apo E.

Linearity

Sample Dilution	Average Percentage of Expected Value	
	Plasma	Serum
1:200	102%	106%
1:400	100%	98%
1:800	96%	97%

Recovery

Standard Added Value	0.05 – 0.5 ug/ml
Recovery %	88 – 108 %
Average Recovery %	98 %

Cross-Reactivity

Species	% Cross Reactivity
Beagle	< 1
Bovine	None
Monkey	< 10 (suggest 1:10 dilution for plasma)
Mouse	< 2 (suggest 1:4 for plasma)
Rat	< 1
Swine	None

Plasma reference value: 40 – 70 µg/ml.

References

- (1) Knott TJ *et al.* (1984) *Nucleic Acids Res.* 12 (9):3909-3915
- (2) Westerterp M *et al.* (2007) *J. Lipid Research* 48:1353-1361
- (3) Jong, MC *et al.* (1999) *Thromb. Vasc. Biol.* 19:472-484
- (4) Kowal, RC *et al.* (1990) *J. Biol. Chem.* 265: 10771-10779
- (5) Weisgraber KH *et al.* (1990) *J. Biol. Chem.* 265:22453-22459
- (6) Gautier T *et al.* (2000) *J. Biol. Chem.* 275:37504-37509
- (7) Dumont L *et al.* (2005) *J. Biol. Chem.* 280(45):38108-38116
- (8) van der Hoogt CC *et al.* (2006) *Biochim. Biophys. Acta* 1761(2):213-220
- (9) Soutar AK *et al.* (1975) *Biochemistry* 14(14):3057-3064
- (10) Sigler GF *et al.* (1976) *Proc. Natl. Acad. Sci. U.S.A.* 73(5):1422-1426
- (11) Berbee JF *et al.* (2006) *FASEB J.* 20(12):2162-2164

Version 2.9R