



AssayMax Mouse Albumin ELISA Kit (Cell Culture Supernatants and Urine)

Catalog No. EMA3201-1

Introduction

Albumin is serum hepatic protein, the most abundant protein in serum and contributes to the maintenance of oncotic pressure as well as to transport of hydrophobic molecules (1). Serum albumin level has been linked in clinical practice to several diseases. Low albumin levels can suggest liver (2), kidney disease (3), inflammation (4), shock (5), and malnutrition (6). On the other hand, high albumin levels usually reflect dehydration (7).

Principal of the Assay

The AssayMax Mouse Albumin ELISA (Enzyme-Linked Immunosorbent Assay) kit is designed for detection of mouse albumin in urine and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique that measures rat albumin in less than 4 hours. A polyclonal antibody specific for mouse albumin has been pre-coated onto a 96-well microplate with removable strips. Albumin in standards and samples is sandwiched by the immobilized polyclonal antibody and biotinylated polyclonal antibody specific for mouse Albumin, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution.

Reagents

- **Mouse Albumin Microplate:** A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against mouse albumin.
- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **Mouse Albumin Standard:** Mouse Albumin in a buffered protein base (1.6 µg, lyophilized).
- **Biotinylated Mouse Albumin Antibody (100x):** A 100-fold concentrated biotinylated polyclonal antibody against mouse Albumin (80 µl).
- **MIX Diluent Concentrate (10x):** A 10-fold concentrated buffered protein base (30 ml).

- **Wash Buffer Concentrate (20x):** A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- **Streptavidin-Peroxidase Conjugate (SP Conjugate):** A 100-fold concentrate (90 µl).
- **Chromogen Substrate:** A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- **Stop Solution:** A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

Storage Condition

- Store kit at 2-8⁰C or -20⁰C upon arrival up to the expiration date.
- Opened MIX Diluent may be stored for up to 1 month at 2-8⁰C. Store reconstituted reagents at -20⁰C or below.
- Opened unused strip wells may return to the foil pouch with the desiccant pack, reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm.
- Pipettes (1-20 µl, 20-200 µl, 200-1000 µl and multiple channel).
- Deionized or distilled reagent grade water.

Sample Collection, Preparation and Storage

- **Cell Culture Supernatants:** Centrifuge cell culture media at 3000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store samples at -20⁰C or below. Avoid repeated freeze-thaw cycles.
- **Urine:** Collect urine using sample pot. Centrifuge samples at 800 x g for 10 minutes and assay. Dilute samples 1:400 into MIX Diluent. Store samples at -20⁰C or below for up to 3 months. Avoid repeated freeze-thaw cycles.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use. If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved.
- **MIX Diluent Concentrate (10x):** Dilute the MIX Diluent 1:10 with reagent grade water. Store for up to 1 month at 2-8⁰C.
- **Standard Curve:** Reconstitute the 1.6 µg of albumin standard with 2 ml of MIX Diluent to generate a stock solution of 800 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. The stock solution can be further dilute 1:2 with MIX Diluent to produce 400 ng/ml standard solution. Prepare duplicate or triplicate standard points by serially diluting the standard solution (400 ng/ml) 1:4 with MIX Diluent to produce 100, 25, 6.25 and 1.56 ng/ml solutions. MIX Diluent serves as the zero standard (0 ng/ml). Any remaining solution should be frozen at -20⁰C.

Standard Point	Dilution	[mAlbumin] (ng/ml)
P1	1 part stock (800 ng/ml) + 1 part MIX Diluent	400.000
P2	1 part P1 + 3 parts MIX Diluent	100.000
P3	1 part P2 + 3 parts MIX Diluent	25.000
P4	1 part P3 + 3 parts MIX Diluent	6.250
P5	1 part P4 + 3 parts MIX Diluent	1.560
P6	MIX Diluent	0.000

- **Biotinylated Mouse Albumin Antibody (100x):** Spin down the antibody briefly and dilute the desired amount of the antibody 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.
- **Wash Buffer Concentrate (20x):** Dilute the Wash Buffer Concentrate 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.

Assay Procedure


- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.
- Wash five times with 200 µl of Wash Buffer manually. Invert the plate each time and decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid. If using a machine wash six times with 300 µl of Wash Buffer and then invert the plate, decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid.
- Add 50 µl of Biotinylated Mouse Albumin Antibody to each well and incubate for one hour.
- Wash a microplate as described above.
- Add 50 µl of Streptavidin-Peroxidase Conjugate to each well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash a microplate as described above.
- Add 50 µl of Chromogen Substrate per well and incubate for about 15 minutes or till the optimal blue color density develops. Gently tap plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm **immediately**. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a Standard Curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

Standard Curve

- The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.



Performance Characteristics

- The minimum detectable dose of Albumin is typically 1 ng/ml.
- Intra-assay and inter-assay coefficients of variation were 4.6% and 7.9% respectively.

Linearity

Sample Dilution	Average Percentage of Expected Value	
	Urine	Cell Culture Supernatant
1:200	96%	100%
1:400	102%	103%
1:800	103%	98%

Recovery

Standard Added Value	2 – 100 ng/ml
Recovery %	94-112 %
Average Recovery %	103 %

Cross-Reactivity

Name	% Cross Reactivity
Bovine Albumin	< 0.01
Human Albumin	< 0.1
Mouse Albumin	100
Rat Albumin	< 0.1
Swine Albumin	< 1

- 10% FBS in culture media will not affect the assay.

References

- (1) Gekle M. (2004) *Annu Rev Physiol.*
- (2) Schindler C *et al.* (1999) *J Hepatol.* 31(6):1132
- (3) Hemmeler MH *et al.* (1997) *Nephrol Dial Transplant.* 12 Suppl 2:57-62
- (4) Sesmilo G *et al.* (2004) *Ann Intern Med.* 133(2):111-22
- (5) Wettstein R *et al.* (2004) *Shock.* 22(4):351-357
- (6) Saito T *et al.* (1991) *Jpn J Surg.* 21(4):402-11
- (7) Strand TA (2004) *Am J Clin Nutr.* 79(3):451-6

Version 5.9R

Related Products

- EA2201-1 AssayMax Human Albumin ELISA Kit (Plasma and Serum samples)
- EA3201-1 AssayMax Human Albumin ELISA Kit (Urine and Cell Culture Supernatant samples)
- EMA2201-1 AssayMax Mouse Albumin ELISA Kit (Plasma and Serum samples)
- ERA2201-1 AssayMax Rat Albumin ELISA Kit (Plasma and Serum samples)
- ERA3201-1 AssayMax Rat Albumin ELISA Kit (Urine and Cell Culture Supernatant samples)
- ETA2201-1 AssayMax Rabbit Albumin ELISA Kit
- EPA3201-1 AssayMax Swine Albumin ELISA Kit (Urine and Cell Culture Supernatant samples)
- EPA2201-1 AssayMax Swine Albumin ELISA Kit (Plasma and Serum samples)