

# AssayMax™ Mouse RBP4 ELISA Kit

Assaypro LLC 3400 Harry S Truman Blvd St. Charles, MO 63301 T (636) 447-9175 F (636) 395-7419 www.assaypro.com

For any questions regarding troubleshooting or performing the assay, please contact our support team at support@assaypro.com.

# **Assay Summary**

**Step 1**. Add 50 μl of Standard or Sample per well. Incubate 2 hours.

**Step 2.** Wash, then add 50  $\mu$ l of Biotinylated Antibody per well. Incubate 1 hour.

**Step 3**. Wash, then add 50  $\mu$ l of SP Conjugate per well. Incubate 30 minutes.

**Step 4.** Wash, then add 50  $\mu$ l of Chromogen Substrate per well. Incubate 30 minutes.

**Step 5.** Add 50  $\mu$ l of Stop Solution per well. Read at 450 nm immediately.

# **Symbol Key**



Consult instructions for use.

# **Assay Template**

12								
11								
10								
6								
∞								
7								
9								
.c								
4								
ю								
2								
1								
	Ą	В	3	Q	3	Ŧ	9	I

# AssayMax™ Mouse Retinol-Binding Protein 4 (RBP4) ELISA Kit

Catalog No. EMR3005-1
Sample insert for reference use only

#### Introduction

Serum retinol-binding protein (RBP4), secreted by the liver and adipocytes, is implicated in systemic insulin resistance. RBP4 transports retinol and circulates in the plasma by binding to the larger transthyretin (TTR) homotetramer, forming a protein complex that reduces renal clearance of RBP4. In insulin-resistant ob/ob mice, urinary fractional excretion of RBP4 was reduced, which is consistent with increased retention, while TTR level was elevated (1). RBP4 is encoded by the RBP4 gene, which maps to chromosome 10q23-q24, and is linked to increased risk for type 2 diabetes in different populations (2-3). Transgenic overexpression of RBP4 or injection of recombinant RBP4 in normal mice causes insulin resistance. Conversely, genetic deletion of RBP4 enhances insulin sensitivity. Increasing serum RBP4 induces hepatic expression of the gluconeogenic enzyme, phosphoenolpyruvate carboxykinase, and impairs insulin signaling in muscle tissue (4). Expression of RBP4 is induced in adipose tissue as a consequence of the decreased glucose transporter GLUT4 expression. Increased serum RBP4 is associated with insulin resistance, type 2 diabetes, and metabolic syndrome such as obesity, glucose intolerance, dyslipidemia, and hypertension (5-6). Plasma RBP4 concentration might be a biomarker of nephropathy and cardiovascular disease in type 2 diabetic subjects (7).

# **Principle of the Assay**

The AssayMax™ Mouse RBP4 ELISA (Enzyme-Linked Immunosorbent Assay) Kit is designed for detection of RBP4 in mouse plasma, serum, urine, cell culture, cell lysate, and tissue samples. This assay employs a quantitative sandwich enzyme immunoassay that measures mouse RBP4 in approximately 4 hours. A polyclonal antibody specific for mouse RBP4 has been pre-coated onto a 96-well microplate with removable strips. RBP4 in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for mouse RBP4, which is recognized by a streptavidin-peroxidase (SP) conjugate. All unbound material is washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

#### **Caution and Warning**

- This product is for Research Use Only and is not intended for use in diagnostic procedures.
- Prepare all reagents (diluent buffer, wash buffer, standard, biotinylated antibody, and SP conjugate), as instructed, prior to running the assay.
- Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this insert. However, the user should determine the optimal dilution factor.
- Spin down the SP conjugate vial and the biotinylated antibody vial before opening and using contents.
- The Stop Solution is an acidic solution.
- The kit should not be used beyond the expiration date.

#### Reagents

- Mouse RBP4 Microplate: A 96-well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody against mouse RBP4.
- Sealing Tapes: Each kit contains 3 precut, pressure sensitive sealing tapes that can be cut to fit the format of the individual assay.
- Mouse RBP4 Standard: Mouse RBP4 in a buffered protein base (60 ng, lyophilized).
- **Biotinylated Mouse RBP4 Antibody (25x):** A 25-fold concentrated biotinylated polyclonal antibody against mouse RBP4 (200 µl).
- MIX Diluent Concentrate (10x): A 10-fold concentrated buffered protein base (30 ml).
- Wash Buffer Concentrate (20x): A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- SP Conjugate (100x): A 100-fold concentrate (80 μl).
- Chromogen Substrate (1x): A stabilized peroxidase chromogen substrate tetramethylbenzidine (7 ml).
- **Stop Solution (1x):** A 0.5 N hydrochloric acid solution to stop the chromogen substrate reaction (11 ml).

#### **Storage Condition**

- Upon arrival, immediately store components of the kit at recommended temperatures up to the expiration date.
- Store SP Conjugate and Biotinylated Antibody at -20°C.
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C.
- Unused microplate wells may be returned to the foil pouch with the desiccant packs and resealed. May be stored for up to 30 days in a vacuum desiccator.

 Store Standard at 2-8°C before reconstituting with Diluent and at -20°C after reconstituting with Diluent.

#### **Other Supplies Required**

- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes (1-20 μl, 20-200 μl, 200-1000 μl, and multiple channel)
- Deionized or distilled reagent grade water

#### Sample Collection, Preparation, and Storage

- Plasma: Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 3000 x g for 10 minutes and collect plasma. An 8000-fold sample dilution is suggested into MIX Diluent; however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 3000 x g for 10 minutes and remove serum. An 8000-fold sample dilution is suggested into MIX Diluent; however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- Urine: Collect urine using sample pot. Centrifuge samples at 800 x g for 10 minutes. A 10-fold sample dilution is suggested into MIX Diluent or within the range of 1x 100x; however, user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- Cell Culture Supernatant: Centrifuge cell culture media at 1500 rpm for 10 minutes at 4°C to remove debris and collect supernatant. If necessary, dilute samples into MIX Diluent; user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -80°C. Avoid repeated freeze-thaw cycles.
- Cell Lysate: Rinse cell with cold PBS and then scrape the cell into a tube with 5 ml of cold PBS and 0.5 M EDTA. Centrifuge suspension at 1500 rpm for 10 minutes at 4°C and aspirate supernatant. Resuspend pellet in ice-cold Lysis Buffer (PBS, 1% Triton X-100, protease inhibitor cocktail). For every 1 x 10<sup>6</sup> cells, add approximately 100 μl of ice-cold Lysis Buffer. Incubate on ice for 60 minutes. Centrifuge at 13000 rpm for 30 minutes at 4°C and collect supernatant. If necessary, dilute samples into MIX Diluent; user should determine optimal dilution factor depending on application needs. The undiluted samples can be stored at -80°C. Avoid repeated freeze-thaw cycles.

• **Tissue:** Extract tissue samples with 0.1 M phosphate-buffered saline (pH 7.4) containing 1% Triton X-100 and centrifuge at 14000 x g for 20 minutes. Collect the supernatant and measure the protein concentration. If necessary, dilute samples into MIX Diluent; user should determine optimal dilution factor depending on application needs. Store remaining extract at -80°C. Avoid repeated freeze-thaw cycles.

Applicable samples may also include biofluids, cell culture, and tissue homogenates. If necessary, user should determine optimal dilution factor depending on application needs.

#### Refer to Dilution Guidelines for further instruction.

	Guidelines for Dilutions of 100-fold or Greater			
	(for reference only; please follow the insert for specific dilution suggested)			
	100x		10000x	
A)	4 μl sample : 396 μl buffer (100x) = 100-fold dilution  Assuming the needed volume is less than or equal to 400 μl.	A) B)	4 μl sample : 396 μl buffer (100x) 4 μl of A : 396 μl buffer (100x) = 10000-fold dilution Assuming the needed volume is less than or equal to 400 μl.	
	1000x		100000x	
A) B)	4 μl sample : 396 μl buffer (100x) 24 μl of A : 216 μl buffer (10x) = 1000-fold dilution  Assuming the needed volume is less than	A) B) C)	4 μl sample : 396 μl buffer (100x) 4 μl of A : 396 μl buffer (100x) 24 μl of B : 216 μl buffer (10x) = 100000-fold dilution Assuming the needed volume is less than	
	or equal to 240 μl.		or equal to 240 μl.	

#### Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- MIX Diluent Concentrate (10x): Dilute the MIX Diluent Concentrate 10-fold with reagent grade water to produce a 1x solution. When diluting the concentrate, make sure to rinse the bottle thoroughly to extract any precipitates left in the bottle. Mix the 1x solution gently until the crystals have completely dissolved. Store for up to 30 days at 2-8°C.
- Mouse RBP4 Standard: Reconstitute the Mouse RBP4 Standard (60 ng) with 1.5 ml of MIX Diluent to generate a 40 ng/ml standard stock solution. Allow the vial to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting from the standard stock solution (40 ng/ml) 2-fold with equal volume of MIX Diluent to produce 20, 10, 5, 2.5, 1.25, and 0.625 ng/ml solutions. MIX Diluent serves as the zero standard (0 ng/ml). Any

remaining stock solution should be stored at -20°C and used within 30 days. Avoid repeated freeze-thaw cycles.

Standard Point	Dilution	[RBP4] (ng/ml)
P1	1 part Standard (40 ng/ml)	40
P2	1 part P1 + 1 part MIX Diluent	20
Р3	1 part P2 + 1 part MIX Diluent	10
P4	1 part P3 + 1 part MIX Diluent	5.0
P5	1 part P4 + 1 part MIX Diluent	2.5
P6	1 part P5 + 1 part MIX Diluent	1.25
P7	1 part P6 + 1 part MIX Diluent	0.625
P8	MIX Diluent	0.0

- Biotinylated Mouse RBP4 Antibody (25x): Spin down the antibody briefly and dilute the desired amount of the antibody 25-fold with MIX Diluent to produce a 1x solution. The undiluted antibody should be stored at -20°C.
- Wash Buffer Concentrate (20x): Dilute the Wash Buffer Concentrate 20fold with reagent grade water to produce a 1x solution. When diluting
  the concentrate, make sure to rinse the bottle thoroughly to extract any
  precipitates left in the bottle. Mix the 1x solution gently until the crystals
  have completely dissolved.
- SP Conjugate (100x): Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 100-fold with MIX Diluent to produce a 1x solution. The undiluted conjugate should be stored at -20°C.

# **Assay Procedure**

- Prepare all reagents, standard solutions, and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-25°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccants inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 µl of Mouse RBP4 Standard or sample to each well. Gently tap
  plate to thoroughly coat the wells. Break any bubbles that may have
  formed. Cover wells with a sealing tape and incubate for 2 hours. Start
  the timer after the last addition.
- Wash the microplate manually or automatically using a microplate washer. Invert the plate and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If washing manually, wash five times with 200 µl of Wash Buffer per well. Invert the

- plate each time and decant the contents; hit 4-5 times on absorbent material to completely remove the liquid. If using a microplate washer, wash six times with 300 µl of Wash Buffer per well; invert the plate and hit 4-5 times on absorbent material to completely remove the liquid.
- Add 50 µl of Biotinylated Mouse RBP4 Antibody to each well. Gently tap
  plate to thoroughly coat the wells. Break any bubbles that may have
  formed. Cover wells with a sealing tape and incubate for 1 hour.
- Wash the microplate as described above.
- Add 50 µl of SP Conjugate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Cover wells with a sealing tape and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.
- Add 50 µl of Chromogen Substrate to each well. Gently tap plate to thoroughly coat the wells. Break any bubbles that may have formed. Incubate in ambient light for 30 minutes or until the optimal blue color density develops.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow. Gently tap plate to ensure thorough mixing. Break any bubbles that may have formed.
- Read the absorbance on a microplate reader at a wavelength of 450 nm immediately. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections.
   Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings.

#### **Data Analysis**

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance (OD) on the y-axis. The best fit line can be determined by regression analysis using log-log or four-parameter logistic curve fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

# **Typical Data**


The typical data is provided for reference only. Individual laboratory
means may vary from the values listed. Variations between laboratories
may be caused by technique differences.

Standard Point	ng/ml	OD	Average OD
P1	40	2.165	2.112
	70	2.059	2.112
P2	20	1.890	1.914
12	20	1.938	1.514
Р3	10	1.486	1.451
гэ	10	1.416	1.431
P4	5.0	0.857	0.876
F ##	5.0	0.895	0.870
P5	2.5	0.443	0.456
ГJ	2.5	0.469	0.430
P6	1.25	0.256	0.265
FU	1.23	0.274	0.203
P7	0.625	0.207	0.200
1 /	0.023	0.193	0.200
P8	0.0	0.143	0.140
го	0.0	0.137	0.140
Sample: Poo	oled Normal	0.525	0.542
Sodium Citrate	Plasma (8000x)	0.561	0.543
Sample: Poo	oled Normal	0.865	0.044
Serum	(8000x)	0.823	0.844

# **Standard Curve**

• The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.

Mouse RBP4 Standard Curve



#### **Performance Characteristics**

- The minimum detectable dose of mouse RBP4 as calculated by 2SD from the mean of a zero standard was established to be 0.44 ng/ml.
- Intra-assay precision was determined by testing three plasma samples twenty times in one assay.
- Inter-assay precision was determined by testing three plasma samples in twenty assays.

	Intra-Assay Precision			Inter	-Assay Pred	ision
Sample	1	2	3	1	2	3
n	20	20	20	20	20	20
CV (%)	3.8%	3.6%	2.2%	10.7%	9.0%	7.9%
Average CV (%)	3.2%			9.2%		

### Recovery

Standard Added Value	1.25 – 20 ng/ml
Recovery %	88 – 116%
Average Recovery %	97%

# Linearity

Plasma and serum samples were serially diluted to test for linearity.

Average Percentage of Expected Value (%)			
Sample Dilution	Plasma	Serum	
4000x	90%	90%	
8000x	97%	99%	
16000x	105%	106%	

# **Cross-Reactivity**

Species	Cross-Reactivity (%)
Canine	None
Bovine	None
Equine	None
Monkey	None
Human	None
Rat	5%
Swine	None
Rabbit	None

• 10% FBS in culture media will not affect the assay.

# Troubleshooting

Issue	Causes	Course of Action
	Use of improper	Check the expiration date listed before use.
	components	<ul> <li>Do not interchange components from different lots.</li> </ul>
		<ul> <li>Check that the correct wash buffer is being used.</li> </ul>
		<ul> <li>Check that all wells are empty after aspiration.</li> </ul>
	Improper wash step	<ul> <li>Check that the microplate washer is dispensing properly.</li> </ul>
		If washing by pipette, check for proper pipetting
드	Cularbina of manager	technique.
Low Precision	Splashing of reagents while loading wells	Pipette properly in a controlled and careful manner.
Pre	Inconsistent volumes	<ul> <li>Pipette properly in a controlled and careful manner.</li> </ul>
3	loaded into wells	Check pipette calibration.
Lo		Check pipette for proper performance.
	Insufficient mixing of	Thoroughly agitate the lyophilized components after
	reagent dilutions	reconstitution.  Thoroughly mix dilutions.
		Check the microplate pouch for proper sealing.
	Improperly sealed	Check that the microplate pouch has no punctures.
	microplate	Check that three desiccants are inside the microplate
		pouch prior to sealing.
	Microplate was left	Each step of the procedure should be performed
la	unattended between	uninterrupted.
igu	steps	
υS	Omission of step	<ul> <li>Consult the provided procedure for complete list of steps.</li> </ul>
ligh	Steps performed in	<ul> <li>Consult the provided procedure for the correct order.</li> </ul>
ı.	incorrect order	
v o sit,	Insufficient amount of reagents added to	Check pipette calibration.     Check pipette for proper performance.
en o	wells	Check pipette for proper performance.
Unexpectedly Low or High Signal Intensity	Wash step was skipped	Consult the provided procedure for all wash steps.
edl	Improper wash buffer	Check that the correct wash buffer is being used.
ğ	Improper reagent	Consult reagent preparation section for the correct
ğ	preparation	dilutions of all reagents.
ne)	Insufficient or	<ul> <li>Consult the provided procedure for correct incubation</li> </ul>
)	prolonged incubation	time.
	periods	
e e		<ul> <li>Sandwich ELISA: If samples generate OD values higher than the highest standard point (P1), dilute samples</li> </ul>
n n		further and repeat the assay.
) r	Non-optimal sample	Competitive ELISA: If samples generate OD values lower
Deficient Standard Curve Fit	dilution	than the highest standard point (P1), dilute samples
t n		further and repeat the assay.
itan Fit		<ul> <li>User should determine the optimal dilution factor for</li> </ul>
±		samples.
ier	Contamination of	A new tip must be used for each addition of different
fic	reagents	samples or reagents during the assay procedure.
De	Contents of wells	Verify that the sealing film is firmly in place before placing
	evaporate	the assay in the incubator or at room temperature.

Improper pipetting	Pipette properly in a controlled and careful manner. Check pipette calibration. Check pipette for proper performance.
Insufficient mixing of reagent dilutions	<ul> <li>Thoroughly agitate the lyophilized components after reconstitution.</li> <li>Thoroughly mix dilutions.</li> </ul>

#### References

- (1) Mody N et al. (2008) Am J Physiol Endocrinol Metab. 294(4):E785-793.
- (2) Meigs JB et al. (2002) Diabetes. 51:833-840.
- (3) Duggirala R et al. (1998) Diabetes. 47(Suppl. 1):A170.
- (4) Yang Q et al. (2005) Nature. 436(7049):356-362.
- (5) Graham TE et al. (2006) N Engl J Med. 354:2552-2563.
- (6) McTernan PG et al. (2007) J Clin Endocrinol Metab. 92:2430-2432.
- (7) Cabre A et al. (2007) J Intern Med. 262(4):496-503.

Version 1.5